Metals In Mount Isa




Airborne Dust and Metallic Particulates Concentrations


This report must not be reproduced except in full. Results or figures from this report must not be used without acknowledgment

Note for reading these tables: These results show HVAS site performance cumulative average reported over an annual period. The results track our performance towards meeting our Environmental Authority (EA) limit, which is indicated by the red line at the top of the graph.


Cumulative Total Suspended Particulates (TSP)




Cumulative Lead (Pb) in Total Suspended Particulates



Cumulative Arsenic (As) in PM10 Particulates



Cumulative Cadmium (Cd) in PM10 Particulates



Last validated HVAS run day: 15/06/2024

Metals In Mount Isa



Analytical Procedures

Total Suspended Particulate (TSP) and Particulate Matter - 10µm (PM10) air quality monitoring is undertaken using High Volume Air Samplers (HVAS). The HVAS units are operated once every six days for a 24-hour period from midnight to midnight (local standard time – UTC/GMT +10 hours) in accordance with AS/NZS 3580. Volume measurements are corrected to standard temperature and pressure (STP) conditions (0°C, 101.3 kPa). Filter papers are collected between runs and analysed by a NATA accredited laboratory. Analysis and validation of data may take up to four weeks from the scheduled run day.

| METHOD | ANALYSING LABORATORY | NATA ACCREDITATION | |
|---|---------------------------------------|--------------------|--|
| In-house method SWI140522 - Sample Collection | Mount Isa Mines Analytical Laboratory | 3978 | |
| AS/NZS 3580.9.3 (2015) | Mount Isa Mines Analytical Laboratory | 3978 | |
| AS/NZS 3580.9.6 (2015) | Mount Isa Mines Analytical Laboratory | 3978 | |
| In house method EA144 - Metals Analysis on Monitoring Filters | ALS Environmental | 825 | |


Calibration and Maintenance

Calibration of the HVAS units is undertaken monthly, in accordance with relevant standards. The table below lists the tolerances used for calibration of the equipment used and the associated measurement uncertainty.

| Equipment | Parameter | CALIBRATION TOLERANCE | UNCERTAINTY ¹ |
|---|---------------------------|---|--------------------------|
| Ecotech HiVol 3000 - TSP PM10 | Temperature | ±1°C of reference meter ² | |
| | Pressure | ±7.5 mmHg of reference meter ² | |
| | Pre-Calibration air flow | ±10% of calculated value | |
| | Post-Calibration air flow | ±1% of calculated value | ±1% |
| | Time | N/A | Negligible |
| Cubis Analytical Balance - MSU224S-100-DU | Mass | Per NATA calibration report | ±0.07mg |

¹Taken from manufacturer specifications

Monitoring Locations



²Reference meters are independently calibrated against a primary source every 6 months by HVAS Manufacturer (Ecotech)